Fiscal Year:	FY 2017	Task Last Updated:	FY 02/05/2018
PI Name:	Strangman, Gary E Ph.D.		
Project Title:	Testing Mechanical Countermeasures for Cephalad Fluid Shifts		
Division Name:	Human Research		
Program/Discipline:	NSBRI		
Program/Discipline Element/Subdiscipline:	NSBRISmart Medical Systems and Technology Team		
Joint Agency Name:		TechPort:	Yes
Human Research Program Elements:	(1) HHC :Human Health Countermeasures		
Human Research Program Risks:	None		
Space Biology Element:	None		
Space Biology Cross-Element Discipline:	None		
Space Biology Special Category:	None		
PI Email:	strang@nmr.mgh.harvard.edu	Fax:	FY
PI Organization Type:	NON-PROFIT	Phone:	617-724-0662
Organization Name:	Massachusetts General Hospital		
PI Address 1:	Department of Psychiatry		
PI Address 2:	149 13th Street, Suite 2651		
PI Web Page:			
City:	Charlestown	State:	MA
Zip Code:	02129-2020	Congressional District:	7
Comments:			
Project Type:	Ground	Solicitation / Funding Source:	2014-15 HERO NNJ14ZSA001N-Crew Health (FLAGSHIP & NSBRI)
Start Date:	06/01/2015	End Date:	05/31/2017
No. of Post Docs:	2	No. of PhD Degrees:	0
No. of PhD Candidates:	0	No. of Master' Degrees:	0
No. of Master's Candidates:	0	No. of Bachelor's Degrees:	0
No. of Bachelor's Candidates:	1	Monitoring Center:	NSBRI
Contact Monitor:		Contact Phone:	
Contact Email:			
Flight Program:			
Flight Assignment:			
Key Personnel Changes/Previous PI:			
COI Name (Institution):	Cohen, Adam M.D. (Massachusetts General Hosp Rosenthal, Eric M.D. (Massachusetts General Hosp Dentinger, Aaron Ph.D. (GE Global Research) Bershad, Eric M.D. (Baylor College of Medicine) Venkatasubba Rao, Chethan M.D. (Baylor College Zhang, Quan Ph.D. (Massachusetts General Hospi Baggish, Aaron M.D., Ph.D. (Massachusetts General	ital) pital) e of Medicine) tal) ral Hospital)	
Grant/Contract No.:	NCC 9-58-SMST04201		
Performance Goal No.:			
Performance Goal Text:			

Task Description:	Current evidence suggests that NASA's spaceflight-associated neuro-opthalmological syndrome (SANS/VIIP) risk is related to an elevation in intracranial pressure (ICP) during spaceflight, either consequent to or aggravated by cephalad fluid shifts in microgravity. From recent data, SANS occurs in >50% of astronauts, to varying degrees, and can lead to long term visual changes. Although its cause is unknown, its importance is high enough to motivate studies of potential countermeasures. The key objectives of this project were therefore: (1) to test and help validate two commercial devices as mechanical countermeasures for cephalad fluid shifts to potentially treat elevated ICP, (2) identify any potential adverse consequences during use or following release of such countermeasures, and (3) optimize deployment procedures for such countermeasures. The Russians currently use Braslet—an elastic thigh band—to help sequester blood in the legs and alleviate symptoms resulting from cephalad fluid shifts. While promising, this device has not been tested as a SANS countermeasure. Lower body negative pressure (LBNP) is an alternative approach, which draws fluid into the legs using vacuum mechanism. Both have drawbacks, however. Braslet devices are custom-built, difficult to obtain, and have limited calibration options. LBNP is typically bulky and hence could only be used at limited times. As an alternative to Braslet, we tested the Kaatsu thigh cuff system. This commercially available system is designed for enhanced muscle training on Earth. In addition, we investigated use of a LymphaPress compression garment configured to provide a vascular resistance for fluid return from the lower body (as opposed to enhanced fluid return for which the device was designed for clinically). In Experiments 1 and 2 we conducted tests using both countermeasures (at different inflation pressures) in healthy subjects undergoing -6 degrees head-down tilt (HDT). We characterized cerebral blood volume and flow, intracoular pressure, structural eye
Rationale for HRP Directed Research:	
	Impact: This project provided the first tests of commercial, user-friendly and safety-tested devices as countermeasures potentially suitable for SANS. Results include: (1) an assessment of both thigh cuffs and a compression garment as a SANS countermeasure, (2) assessment of the influence of these devices on cerebrovascular and ocular parameters, (3) parameterization of countermeasure deployment and rebound effects on multiple physiological variables, and (4) information regarding optimized deployment of the countermeasures.
Research Impact/Earth Benefits:	Earth Benefits: Currently, there are few treatment methods for elevated ICP, which affects patients with traumatic brain injury, stroke, hydrocephalus, and cancer patients. None of the current methods involve non-invasive mechanical devices—instead focusing on surgical procedures or medications. This work therefore has the potential to identify one or more countermeasures and/or protocols—within a novel class of countermeasures—that could be used to help manage intracranial fluids and pressure. Since these approaches do not require drugs, they avoid the potential side effects, drug-drug interactions, or longer-lasting effects that often come from medication use.
Task Progress:	Current evidence suggests that NASA's spaceflight associated neuro-opthalmological syndrome (SANS/VIIP) risk may be related to an elevation in intracranial pressure (ICP) during spaceflight compared to the upright position on Earth, either consequent to or aggravated by cephalad fluid shifts in microgravity. The key objectives of this project were: (1) to test and assess two commercial devices as mechanical countermeasures for cephalad fluid shifts and modify cerebral and/or ocular parameters, (2) to identify any potential adverse consequences during use or following release of such countermeasures, and (3) to optimize deployment procedures for such countermeasures. Specifically, this project sought to test the Kaatsu thigh cuff muscle training system, as an alternative to the Braslet, and a reconfigured Lymphapress system, originally designed for fluid management in edema patients. In year 2 of the project we initiated and completed three human experiments. In Experiment 1, we tested n=18 healthy subjects during a -6 degree head-down tilt protocol. Subjects were monitored during two 3-hour sessions (one for each countermeasure, randomized order) consisting of 20-min periods in each of the following orientations, in sequence: +50 degree HUT (head up tilt), supine, -6 degree HDT, -6 degree HDT during countermeasure deployment, -6 degree HDT post-deployment of the countermeasure, supine, and briefly again at 50 degree HUT. All subjects were tested with Kaatsu at 180 SKU deployment pressure and Lymphapress at 34-36 mmHg pressures, in counterbalanced orders. Experiment 2 tested n=12 healthy subjects identically as in Experiment 1, however, with higher countermeasure pressures: Kaatsu at a deployment pressure of 300 SKU and Lymphapress at pressures of 43-45 mmHg. In Experiment 3, the Kaatsu system was tested in n=5 patients in the NeuroICU in collaboration with Dr. Eric Bershad with invasive ICP monitoring for 20 min Kaatsu deployment, and 20 min post-deployment at 250 SKU. Neither of the investigated mechanical cou
Bibliography Type:	Description: (Last Updated: 02/05/2025)
Articles in Peer-reviewed Journals	Strangman GE, Ivkovic V, Zhang Q. "Wearable brain imaging with multi-modal physiological recording." J Appl Physiol (1985). 2018 Mar 1;124(3):564-72. Epub 2017 Jul 13. <u>https://doi.org/10.1152/japplphysiol.00297.2017</u> ; PubMed <u>PMID: 28705994</u> [reported originally in July 2017 as "Epub ahead of print"], Mar-2018