Fiscal Year:	FY 2012	Task Last Updated:	FY 03/08/2012
PI Name:	Borak, Thomas B. Ph.D.		
Project Title:	Lunar EVA Dosimetry: Design of a Radiation I	Dosimeter for Astronauts During L	unar Extravehicular Activities
Division Name:	Human Research		
Program/Discipline:	NSBRI		
Program/Discipline Element/Subdiscipline:	NSBRIRadiation Effects Team		
Joint Agency Name:		TechPort:	Yes
Human Research Program Elements:	(1) SR :Space Radiation		
Human Research Program Risks:	(1) ARS:Risk of Acute Radiation Syndromes D	Oue to Solar Particle Events (SPEs)	
Space Biology Element:	None		
Space Biology Cross-Element Discipline:	None		
Space Biology Special Category:	None		
PI Email:	tborak@colostate.edu	Fax:	FY 970 491 0623
PI Organization Type:	UNIVERSITY	Phone:	970-491-6450
Organization Name:	Colorado State University		
PI Address 1:	Environmental & Radiological Health Sciences		
PI Address 2:	1618 Campus Delivery		
PI Web Page:			
City:	Fort Collins	State:	СО
Zip Code:	80523-1618	Congressional District:	4
Comments:			
Project Type:	Ground	Solicitation / Funding Source:	2006 Radiation Biology NNJ06ZSA001N
Start Date:	11/01/2007	End Date:	10/31/2011
No. of Post Docs:	0	No. of PhD Degrees:	0
No. of PhD Candidates:	1	No. of Master' Degrees:	2
No. of Master's Candidates:	1	No. of Bachelor's Degrees:	0
No. of Bachelor's Candidates:	0	Monitoring Center:	NSBRI
Contact Monitor:		Contact Phone:	
Contact Email:			
Flight Program:			
Flight Assignment:	NOTE: title changed per NSBRI (12/08)		
Key Personnel Changes/Previous PI:			
COI Name (Institution):	Braby, Leslie (Texas Engineering Experiment	t Station)	
Grant/Contract No.:	NCC 9-58-RE01301		
Performance Goal No.:			
Performance Goal Text:			
	Task 1: Design, Fabrication, and Testing Tissue The purpose of this task was to design, build, an (TEPC) that would satisfy the basic specification (extravehicular activities) and as area monitors anode with recessed guard ring insulators to sha mm and the wall thickness is 3 mm for a total d thickness of 0.5 mm were designed and gold pla mass spectrometer was assembled to measure v	e Equivalent Proportional Counters nd assemble a prototype Tissue Equipons outlined by NASA for a dosime in space craft and habitats. The sphape the electric field near the poles. liameter of 24 mm (~1 inch). Alum ated to maintain electrical conduction acuum leaks for the assembled detert	(TEPC) Detectors ivalent Proportional Counters ter for astronauts during lunar EVAs erical TEPC is based on a single-wire The diameter of the gas cavity is 18 inum vacuum chambers with a shell vity. A system using a high sensitivity ectors with high special specificity.

1.1

- - - - -

	We have been using a version of the software package LOREN1Z 3D to model the electric field inside a spherical detector with a linear collector. This uses special modeling techniques based on the Boundary Element Method to make the solution of these very challenging problems a simple matter. The geometry of the problem can be created with the geometric modeler built into the electric field solvers or can be imported from any of the major CAD (computer-aided design) vendors. More importantly, the geometry can be changed parametrically to optimize a design for robustness, weight, size and, of course, cost. We have fabricated seven versions of the TEPC and Vacuum Chamber. Three versions with spherical detector and single wire anode operated with the wall at high voltage and the anode at ground were delivered for electronics
Task Description:	development and testing. Two versions with a spherical detector and single wire anode operated with the wall at ground and the anode at high voltage have been used for TEPC development and comparative analysis. Two versions of a detector using a new hybrid design with a parallel wire grid surrounding the anode we designed and built. The objective of this design was to form a virtual cylindrical geometry around the anode wet designed and built. The objective of this design was to form a virtual cylindrical geometry around the anode with that would improve the spatial resolution of the TEPC without distorting the signals required for microdosimetry applications. The single and multi wire detectors with grounded anodes were exposed to Neutrons from a PuBe source at Colorado State University and high energy charged particle beams at the Heavy Ion Medical Accelerator in Chiba (HIMAC) synchrotron in Chiba Japan and the NASA Space Radiation Laboratory (NSRL) at Brookhaven National Laboratory. This included the following ions and energies: 56Fe (380 MeV/amu), 18Ar (300 MeV/amu), 12C (200 MeV/amu), and 1H (230 MeV/amu). Measurements were taken at several angles of incidence to determine the angular response of the detector. These results were compared with similar measurements using a commercial TEPC with a Rossi design that has a helical grid surrounding the anode to provide a uniform angular response. Analysis of detector response using digital signal processing was initiated. This system replaced a preamplifier and shaping amplifier with a programmable logic device (PLD) that captures the output signal from the TEPC and digitizes the amplitude in 10 nsec intervals.
	Task 2: Modeling Detector Response
	The objective of this task is to determine the response of the TEPC under ambient conditions and during solar particle events (SPE) on the lunar surface. Computations using the Monte Carlo Code PHITS have been made to determine the energy deposition in the TEPC using protons with an energy spectrum from a SPE in October 2003. These data were compared with the dose that would be delivered to the skin beneath a space suit with an areal density of 0.4 g/cm2. It is clear that a stainless steel vacuum chamber in Mod 1 needs to be replaced with lighter and thinner materials. These results will be important in determining what additional modifications will be necessary to achieve the design goal for real time measurements to the skin and blood forming organs (BFO).
	Task 3: Modeling the Variance-Covariance Method
	The original proposal for the EVA dosimeter was based on the concept of having two independent proportional counters that would be used to obtain estimates of dose, D, and a quality factor, Q, based on estimating using the variance-covariance method. It was recognized that because of size limitations, the proportional counters would have to be located too close to one another to satisfy the condition that a single particle could not intercept both detectors. The additional constraint that one of the detectors must measure the dose at the skin surface and the other at a depth corresponding to the blood forming organs, makes the original variance-covariance method with paired detectors impractical. We have developed a method using a single detector in a variance-covariance scheme. The concepts are based on collecting the charge, q(i), in a single TEPC for n successive time intervals, i. The method proposed by Borak at Colorado State University (CSU) separates the data set into two groups of n/2 entries of values for q(i) based on odd and even indices. The n/2 pairs of data (odd and even) are used to obtain the covariance and each of the two sets of n/2 values (odd or even) to estimate a variance. Monte Carlo codes have been written to test the algorithmic using microdosimetric spectra obtained from measurements in Task 1. The tests indicated that if the change in dose rate and dose averaged lineal energy for estimating quality factors. The analysis also indicated that the estimate of dose mean lineal energy for high energy heavy ions (HZE) and recoil protons from PuBe neutrons converged to the correct value when the number of intervals exceeded 100 and the width of each interval was selected such that the mean number of events in each time interval (i) was less than 30.
Rationale for HRP Directed Research	:
Research Impact/Earth Benefits:	This type of dosimeter has additional applications for first responders to nuclear accidents or terrorist events. It can also provide real time dosimetry during high altitude aviation and commercial space flight, diagnostic, and therapeutic medical procedures such as proton and carbon ion radiation therapy, and surveillance activities associated with homeland security and nuclear non proliferation. It can also serve an a area monitor with live-time capabilities that provide dose rate as well as estimates of quality factors for radiation workers as well as the general public.
	Task 1: Design and Fabrication of Tissue Equivalent Proportional Counters Three Detector systems have been delivered to Texas A&M University (TAMU) for installation of the preamplifier and on to NASA Ames Research Center for data acquisition and analysis hardware. The spherical TEPC is based on a single-wire anode with recessed guard ring insulators to shape the electric field near the poles. The diameter of the gas cavity is 18 mm and the wall thickness is 3 mm for a total diameter of 24 mm (~1 inch). A gold plated aluminum vacuum chamber designed to accommodate the TEPC and pre amplifier has been fabricated and leak tested. The hemispherical dome and vacuum chamber wall surrounding the TEPC has a thickness of 0.5 mm. The units were assembled and leak tested using a He vacuum leak test system.
	Task 2: Determine the response of the TEPC to low energy protons expected during a SPE
Task Progress:	The objective of this task is to determine the response of the TEPC under ambient conditions and during SPE events on the lunar surface. Computations using the Monte Carlo Code PHITS have been made to determine energy deposition in the TEPC using protons with an energy spectrum from a SPE in October 2003. These data were compared with the dose that would be delivered to the skin beneath a space suit with an areal density of 0.4 g/cm2. These results will be important in determining what modifications will be necessary to achieve the design goal that the EVA dosimeter provides real time measurements to the skin and BFO.

e original variance/covariance method requires two independent detectors to measure dose and dose averaged lineal rgy that is used to obtain a radiation quality factor for the incident particles. We have been developing a method ng a single detector in a variance-covariance scheme The concepts are based on collecting the charge, q(i), in a single PC for n successive intervals of (i) with the condition that the change in dose rate is very small between (i) and (i+1). e method simulates two detectors by separating the data set into two groups of n/2 entries of values for q(i) based on and even indices. The n/2 pairs of data (odd and even) are used to obtain the covariance between odd and even asurements and each of the two sets of n/2 values (odd or even) to estimate a variance. Monte Carlo codes have been reloped to test the algorithms in terms of the width of each interval of (i) and the number of intervals, n, that need to collected. The estimate of dose mean lineal energy converged to the correct value when n was greater than 100. The th of the interval, specified by the mean number of events in each interval (i), yielded consistent results from a mean nber of events from 1 to 40 for all three distributions. Tests have been made to determine the effectiveness of dose e changes between charge collection intervals.
scription: (Last Updated: 03/20/2019)
en C, Borak TB, Tsujii H, Nickoloff JA. "Heavy charged particle radiobiology: using enhanced biological ectiveness and improved beam focusing to advance cancer therapy." Mutat Res. 2011 Jun 3;711(1-2):150-7. Epub 11 Mar 3. Review. PubMed <u>PMID: 21376738</u> ; <u>http://dx.doi.org/10.1016/j.mrfmmm.2011.02.012</u> , Jun-2011
rak TB, Chapman PL. "Description and verification of an algorithm for obtaining microdosimetric quantities for h-LET radiation using a single TEPC without pulse height analysis." Radiat Res. 2014 Oct;182(4):396-407. b://dx.doi.org/10.1667/RR13266.1; PubMed <u>PMID: 25211132</u> , Oct-2014
aume T, Braby LA, Borak TB, Lusby T, Warner DW, Perez-Nunez D. "Compact tissue-equivalent proportional inter for deep space human missions." Health Phys. 2015 Oct;109(4):277-83. PubMed <u>PMID: 26313585</u> ; PubMed ntral <u>PMCID: PMC4554228</u> , Oct-2015