

<b>Fiscal Year:</b>	FY 2012	<b>Task Last Updated:</b>	FY 10/19/2012
<b>PI Name:</b>	Ritter, James A Ph.D.		
<b>Project Title:</b>	Development of Pressure Swing Adsorption Technology for Spaceflight Medical Oxygen Concentrators		
<b>Division Name:</b>	Human Research		
<b>Program/Discipline:</b>	NSBRI		
<b>Program/Discipline--Element/Subdiscipline:</b>	NSBRI--Smart Medical Systems and Technology Team		
<b>Joint Agency Name:</b>	<b>TechPort:</b>	Yes	
<b>Human Research Program Elements:</b>	(1) <b>ExMC</b> :Exploration Medical Capabilities		
<b>Human Research Program Risks:</b>	(1) <b>ExMC</b> :Risk of Unacceptable Health and Mission Outcomes Due to Limitations of In-flight Medical Capabilities (IRP Rev E)		
<b>Space Biology Element:</b>	None		
<b>Space Biology Cross-Element Discipline:</b>	None		
<b>Space Biology Special Category:</b>	None		
<b>PI Email:</b>	<a href="mailto:ritter@enr.sc.edu">ritter@enr.sc.edu</a>	<b>Fax:</b>	FY
<b>PI Organization Type:</b>	UNIVERSITY	<b>Phone:</b>	803-777-3590
<b>Organization Name:</b>	University of South Carolina		
<b>PI Address 1:</b>	3C07 Swearingen Engineering Center		
<b>PI Address 2:</b>	Department of Chemical Engineering		
<b>PI Web Page:</b>			
<b>City:</b>	Columbia	<b>State:</b>	SC
<b>Zip Code:</b>	29208-4101	<b>Congressional District:</b>	6
<b>Comments:</b>			
<b>Project Type:</b>	GROUND	<b>Solicitation / Funding Source:</b>	2008 Crew Health NNJ08ZSA002N
<b>Start Date:</b>	09/01/2009	<b>End Date:</b>	08/31/2013
<b>No. of Post Docs:</b>	1	<b>No. of PhD Degrees:</b>	1
<b>No. of PhD Candidates:</b>	4	<b>No. of Master' Degrees:</b>	0
<b>No. of Master's Candidates:</b>	1	<b>No. of Bachelor's Degrees:</b>	2
<b>No. of Bachelor's Candidates:</b>	2	<b>Monitoring Center:</b>	NSBRI
<b>Contact Monitor:</b>	<b>Contact Phone:</b>		
<b>Contact Email:</b>			
<b>Flight Program:</b>			
<b>Flight Assignment:</b>			
<b>Key Personnel Changes/Previous PI:</b>			
<b>COI Name (Institution):</b>	Knox, James ( NASA Marshall Space Flight Center ) Edwards, Paul ( SeQual Technologies ) LeVan, Douglas ( Vanderbilt University )		
<b>Grant/Contract No.:</b>	NCC 9-58-SMST02002		
<b>Performance Goal No.:</b>			
<b>Performance Goal Text:</b>			

**Task Description:**

A source of medical oxygen will be needed at some point to keep an astronaut alive during a space mission. To meet this need, the ideal oxygen source would be a light, compact unit that uses minimal electricity, and can supply oxygen continuously for many days. No current technology meets these requirements. Traditional compressed-oxygen cylinders provide a limited amount of oxygen in a heavy, inconvenient package and are not suited for space missions. Oxygen concentrators, which extract oxygen from air using electricity, can eliminate the obvious problems with cylinder storage in space. These kinds of medical oxygen concentrators are already used in residential and military applications. However, existing systems are too big, use too much power, and are too heavy to be carried into space. For example, a unit that can produce oxygen continuously at 4 LPM, weigh less than 7 pounds and use less than 100 Watts of electric power requires a two-fold reduction in weight and power consumption, compared with the most advanced oxygen concentrators now in production by SeQual. As proposed herein, this requirement may be met by combining new air compressor designs with advances in Pressure Swing Adsorption (PSA) technology. SeQual and the team of researchers from the University of South Carolina, Vanderbilt University and the Marshall Space Flight Center are uniquely positioned to achieve this next level of performance.

To determine whether the proposed technology advances are indeed possible, during the second year of this four year project, the four teams of researchers have been busy carrying out extensive mathematical modeling studies (USC), measuring equilibrium and kinetic parameters for the modeling effort (VU), performing carefully planned experiments with an Eclipse medical oxygen system modified for testing at the bench scale (SeQual), and gearing up for testing an Eclipse medical oxygen system under different environmental conditions (MSFC). Results from numerous experiments were used successfully to validate USC's Dynamic Adsorption Process Simulator (DAPS). In particular, DAPS was specially modified and calibrated against a SeQual PSA module under controlled conditions with a decoupled compressor, and the process performance was analyzed with respect to cycle speed, temperature and high to low pressure ratio. Once validated, DAPS simulations focused on varying certain key process parameters to arrive at optimized PSA cycle designs. The learning from the design effort was implemented into a modified PSA module design operating a new PSA cycle, larger feed/exhaust ports, a backfill step, and larger recycle and purge ports. The new PSA module, associated compressor and other components were fabricated and assembled on a breadboard. The breadboard was connected to instrumentation and tested. The new PSA design successfully delivered 4 lpm of product in about an 8 lb assembly with a compressor shaft power of 130 Watts. This was a significant outcome, especially since the new PSA design was based entirely on predictions from the DAPS. Overall, in the first two years of this four year project, this program is ahead of schedule and definitely on track for improving even further the efficiency of the PSA separation, with the project potentially culminating in a breadboard system that will supply 4 LPM of oxygen, weigh 7.2 lbs, require 106 Watts, and satisfy any new constraints imposed by NASA.

During year 3 the task outline presented in the original proposal was followed. In this way, carefully planned experiments carried out by the folks at SeQual were used to calibrate and further validate DAPS at USC. This was done in an attempt to further improve the performance of the PSA module and to understand the effects of potential process changes on its performance. These results with DAPS will be obtained in year 4. SeQual also continued to develop their medical oxygen system based, in part, on the simulation results obtained from DAPS. These developments included breadboard testing, further optimization of bed and PSA cycle design, new prototype subcomponent detailed design and fabrication, new prototype preliminary tests, and improving on their process design and mechanical design capabilities. The team at Vanderbilt is continuing to measure and provide equilibrium and mass transfer properties for adsorbate-adsorbent pairs of interest to NASA adsorption technology. In addition, the entire medical oxygen system is being evaluated based on new constraints imposed by NASA. During year 3, testing in a vacuum chamber with an Eclipse medical oxygen system has also been underway at the MSFC to determine how it performs under International Space Station (ISS) environmental conditions.

Year 4 will continue to follow the task outline presented in the original proposal. In this way, based on the best predictions from DAPS, a breadboard system will be built at SeQual and tested there, by the folks at the MSFC and by a team at Glenn. The intent of the team at Glen is to down-select from the medical oxygen systems under their consideration.

**Rationale for HRP Directed Research:****Research Impact/Earth Benefits:**

A major expectation of the research is the development of smaller medical oxygen concentrators, which will be of benefit not only for space flight but also for medical patients on Earth in need of oxygen enriched air.

**Task Progress:**

There are 8 tasks associated with this project. These tasks are listed below. All are on or ahead of schedule. In the year 1, Tasks 1, 2, and 6 were initiated. In the year 2, in addition, Tasks 3 and 4 were initiated, and Task 5 was initiated ahead of schedule. In year 3 Tasks 1-6 were all underway. Progress has been made for each of these tasks. More detail is provided below.

Task 1. Refine Model Parameters: Vanderbilt has been working with USC to update the dynamic cyclic adsorption process simulator (DAPS) with the most up to date thermodynamic and kinetic parameters.

Task 2. Validate DAPS: USC has been working with SeQual to obtain system dimensions, operating conditions and extensive experimental performance data of SeQual's Eclipse system and then using it to calibrate and validate DAPS. Significant progress has been made with respect to DAPS quantitatively predicting the performance of the Eclipse system.

Task 3. Optimize and Understand the SeQual PSA Cycle: Using the refined and validated DAPS, USC, with input from SeQual, have been carrying out extensive parametric studies of SeQual's PSA cycle to determine if it is possible to improve oxygen recovery, productivity or both while maintaining the oxygen purity and without redesigning the PSA module. There have been some key findings with DAPS. Some of these findings were recently verified experimentally by SeQual.

Task 4. Examine Alternative PSA Cycles: Using the refined DAPS, USC, with input from SeQual, from SeQual, have been exploring new PSA cycle designs and cycle schedules to determine if it might be possible to improve the oxygen recovery, productivity or both while maintaining the oxygen purity by redesigning the PSA module.

Task 5. Redesign and Build Improved PSA Module: Based on DAPS predictions, SeQual designed a new PSA module that successfully delivered 4 lpm of product in about an 8 lb assembly with a compressor shaft power of 130 Watts.

Task 6. Define Compressor Specifications and Build Feasibility Prototype for 4 LPM System: SeQual has an operating compressor suitable for a 3 LPM oxygen PSA system through a different funding source. Specifications and requirements have been identified and a feasibility prototype is being built to provide sufficient pressure and vacuum to supply the 4 LPM system.

Task 7. Assemble and Test Breadboard Systems: SeQual will assemble at least two breadboard demonstration systems that incorporate the new PSA module with the existing reciprocating or possibly a redesigned compressor. These breadboard systems will be tested by SeQual, the MSFC and Glen and used to determine new weight and performance targets.

Task 8. Verify DAPS Predictions of New PSA Modules: Using the refined cyclic adsorption process simulator, USC will carry out studies of redesigned systems or new prototypes to verify the simulation results, to determine optimum operating conditions, and to understand the performance limits of the new systems.

<b>Bibliography Type:</b>	Description: (Last Updated: 08/28/2015)
<b>Articles in Peer-reviewed Journals</b>	Bhadra SJ, Ebner AD, Ritter JA. "On the use of the dual process Langmuir model for predicting unary and binary isosteric heats of adsorption." <i>Langmuir</i> . 2012 May 1;28(17):6935-41. Epub 2012 Apr 18. <a href="http://dx.doi.org/">http://dx.doi.org/</a> ; PubMed PMID: 22480343, May-2012
<b>Articles in Peer-reviewed Journals</b>	Giesy TJ, Wang Y, LeVan MD. "Measurement of mass transfer rates in adsorbents: new combined-technique frequency response apparatus and application to CO2 in 13x zeolite." <i>Industrial &amp; Engineering Chemistry Research</i> . 2012, Sep 5;51(35):11509-17. <a href="http://dx.doi.org/">http://dx.doi.org/</a> , Sep-2012
<b>Awards</b>	Ritter JA. "Named a Fellow of the American Chemical Society, July 2012." Jul-2012
<b>Awards</b>	Ritter JA. "Recipient of the 2012 USC Educational Foundation Research Award for Science, Mathematics, and Engineering, May 2012." May-2012