

<b>Fiscal Year:</b>	FY 2013	<b>Task Last Updated:</b>	FY 01/30/2014
<b>PI Name:</b>	Zoldak, John		
<b>Project Title:</b>	Medical Consumables Tracking-GRC		
<b>Division Name:</b>	Human Research		
<b>Program/Discipline:</b>	HUMAN RESEARCH		
<b>Program/Discipline-- Element/Subdiscipline:</b>			
<b>Joint Agency Name:</b>	<b>TechPort:</b>	Yes	
<b>Human Research Program Elements:</b>	(1) <b>ExMC</b> :Exploration Medical Capabilities		
<b>Human Research Program Risks:</b>	(1) <b>Medical</b> :Risk of Adverse Health Outcomes & Decrements in Performance due to Inflight Medical Conditions (IRP Rev I) (2) <b>Stability</b> :Risk of Ineffective or Toxic Medications Due to Long Term Storage (IRP Rev F)		
<b>Space Biology Element:</b>	None		
<b>Space Biology Cross-Element Discipline:</b>	None		
<b>Space Biology Special Category:</b>	None		
<b>PI Email:</b>	<a href="mailto:zoldakj@zin-tech.com">zoldakj@zin-tech.com</a>	<b>Fax:</b>	FY
<b>PI Organization Type:</b>	INDUSTRY	<b>Phone:</b>	440.625.2334
<b>Organization Name:</b>	Zin Technologies, Inc.		
<b>PI Address 1:</b>	GRC-MS00		
<b>PI Address 2:</b>	6745 Engle Road		
<b>PI Web Page:</b>			
<b>City:</b>	Cleveland	<b>State:</b>	OH
<b>Zip Code:</b>	44130-7994	<b>Congressional District:</b>	16
<b>Comments:</b>			
<b>Project Type:</b>	FLIGHT,GROUND	<b>Solicitation:</b>	Directed Research
<b>Start Date:</b>	06/01/2009	<b>End Date:</b>	06/30/2017
<b>No. of Post Docs:</b>	0	<b>No. of PhD Degrees:</b>	0
<b>No. of PhD Candidates:</b>	0	<b>No. of Master' Degrees:</b>	0
<b>No. of Master's Candidates:</b>	0	<b>No. of Bachelor's Degrees:</b>	0
<b>No. of Bachelor's Candidates:</b>	0	<b>Monitoring Center:</b>	NASA JSC
<b>Contact Monitor:</b>	Antonsen, Erik	<b>Contact Phone:</b>	281.483.4961
<b>Contact Email:</b>	<a href="mailto:erik.lantonsen@nasa.gov">erik.lantonsen@nasa.gov</a>		
<b>Flight Program:</b>	ISS		
<b>Flight Assignment:</b>	NOTE: End date changed to 6/30/2017 per PI (Ed., 1/25/17) NOTE: Gap changes per IRP Rev E (Ed., 1/30/14) NOTE: End date is now 09/30/2016 per HRP Master Task List dated 12/28/2012 (Ed., 3/14/2013)		
<b>Key Personnel Changes/Previous PI:</b>			
<b>COI Name (Institution):</b>			
<b>Grant/Contract No.:</b>	Directed Research		
<b>Performance Goal No.:</b>			
<b>Performance Goal Text:</b>			

<b>Task Description:</b>	<p>This task will identify current practices and develop controls, processes, and technical solutions to accurately track the inventory of medical consumables. When shuttle and International Space Station (ISS) medical kits are returned to Earth, fewer medical consumables remain in the medical kits than would be expected based on reported use by the astronauts. This is significant because the possibility exists that exploration missions could be undersupplied and run the risk of not being able to treat an ill or injured crewmember, particularly given the small volume available for the medical kits. This task will identify current practices and develop controls, processes, and technical solutions to accurately track the inventory of medical consumables.</p> <p>The Medical Consumables Tracking (MCT) system will utilize an electronic identification system comprised of a reader/scanner/interrogator and a transponder. The system will address insufficient tracking of medication and medical consumables usage on the ISS. An electronic tag will be placed on each package. Periodically the system will be powered and contents within the Resource Supply Rack (RSR) storage locker will be inventoried. The information in the MCT database will be downlinked to the Health Management System (HMS) Inventory Tracking Tool (HITT), which contains medical inventory information to be accessed on the ground.</p> <p>Results to date (Preliminary Design Review level) indicate that using an RFID tag (electronic tag) to read a highly dense population of medical items (including liquids, pharmaceuticals of a dielectric nature, and metal wrapped packaging) and achieve the minimum accuracy is feasible.</p> <p>Specific aims:</p> <ol style="list-style-type: none"> <li>1. Track medical consumables and medications accessed with an accuracy of 95% or better</li> <li>2. No additional scheduled crew time to access a medication or medical consumable</li> <li>3. Work in a microgravity environment</li> <li>4. Work in space vehicles that have electronically noisy environments</li> <li>5. Encrypted data transfer</li> <li>6. To know how much of a particular medication is available</li> <li>7. Read the electronic ID tagged items inside of Convenience Medications Pack</li> <li>8. Have the ability to scale; to track small quantities or large quantities of medicines and medical consumables.</li> </ol>
<b>Rationale for HRP Directed Research:</b>	<p>This research is directed because it contains highly constrained research, which requires focused and constrained data gathering and analysis that is more appropriately obtained through a non-competitive proposal.</p>
<b>Research Impact/Earth Benefits:</b>	<p>MCT will develop a new antenna system that will allow a high density of tags in a low volume. This technology can be used in hospitals and pharmacies.</p>
<b>Task Progress:</b>	<p>The Medical Consumables Tracking (MCT) system successfully completed their critical design review in September 2013. MCT was selected for flight via the Human Research Program (HRP) select for flight process.</p>
<b>Bibliography Type:</b>	<p>Description: (Last Updated: )</p>